



CERTIFICATE OF ACCREDITATION

ANSI-ASQ National Accreditation Board

500 Montgomery Street, Suite 625, Alexandria, VA 22314, 877-344-3044

This is to certify that

Scale Systems, Inc.
3905 Steve Reynolds Blvd Suite 100
Norcross, GA 30093

has been assessed by ANAB
and meets the requirements of international standard

ISO/IEC 17025:2005

while demonstrating technical competence in the field of

CALIBRATION

Refer to the accompanying Scope of Accreditation for information regarding the types of calibrations to which this accreditation applies.

L1182-1
Certificate Number


ANAB Approval

Certificate Valid: 05/23/2017-05/24/2020
Version No. 001 Issued: 05/23/2017



This laboratory is accredited in accordance with the recognized International Standard ISO/IEC 17025:2005. This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory quality management system (refer to joint ISO-ILAC-IAF Communiqué dated April 2017).

SCOPE OF ACCREDITATION TO ISO/IEC 17025:2005

Scale Systems, Inc.

3905 Steve Reynolds Blvd., Suite 100
 Norcross, GA 30093
 Mac Harrison
 706-790-4546

CALIBRATION

Valid to: **May 24, 2020**

Certificate Number: **L1182-1**

Mass

Parameter / Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method and/or Equipment
Analytical Balances (0.000 1 g Resolution)	(0 to 160) g	0.51 mg	ASTM E617 Class 1 Weights and NIST Handbook 44 utilized for the calibration of the Weighing System
(0.01 g Resolution)	(0 to 4 000) g	12 mg	
Lab Balances (0.1 g Resolution)	(0 to 5 100) g	0.08 g	ASTM E617 Class 2 Weights and NIST Handbook 44 utilized for the calibration of the Weighing System
(1 g Resolution)	(0 to 15) kg	0.8 g	
Industrial Scales ² (0.01 lb Resolution)	(0 to 100) lb	0.015 lb	NIST Class F Weights and NIST Handbook 44 utilized for the calibration of the Weighing System
(0.05 lb Resolution)	(0 to 150) lb	0.044 lb	
(1 lb Resolution)	(0 to 5 000) lb	1.1 lb	
(2 lb Resolution)	(0 to 10 000) lb	2.1 lb	
(5 lb Resolution)	(0 to 40 000) lb	7 lb	
(10 lb Resolution)	(0 to 100 000) lb	17 lb	
Vehicle Scales (20 lb Resolution)	(0 to 200 000) lb	35 lb	NIST Class F Weights and NIST Handbook 44 utilized for the calibration of the Weighing System

Calibration and Measurement Capability (CMC) is expressed in terms of the measurement parameter, measurement range, expanded uncertainty of measurement and reference standard, method, and/or equipment. The expanded uncertainty of measurement is expressed as the standard uncertainty of the measurement multiplied by a coverage factor of 2 ($k=2$), corresponding to a confidence level of approximately 95%.

Notes:

1. On-site calibration service is available for this parameter, since on-site conditions are typically more variable than those in the laboratory, larger measurement uncertainties are expected on-site than what is reported on the accredited scope
2. This scope is formatted as part of a single document including Certificate of Accreditation No. L1182-1.

